

Verification of Conformity

Certificate No.: CISRC240411070-R1

Applicant: Shenzhen Kenqing Technology Co.,Ltd

Address: A1, 1st floor, block A, Shanshui building, Nanshan cloud Valley
Innovation Industrial Park, Liuxian Avenue, Nanshan District,
Shenzhen

Product Name: Lumbar Support Robot

Trade Mark : VIGX、Exopace、Kenqing

Model/Type reference: Ant-Y20

Listed Model(s) : Ant-Y10、 Ant-W1 MIX

Issue date: May 08, 2024

In accordance with the following applicable Directives:

Radio Equipment Directive (RED) 2014/53/EU

The equipment, as described herewith, was tested pursuant to applicable test procedure and complies with the requirements of:

Essential Requirement		Applied Standards	Test Report No.	Result
Art. 3.1a	Health	EN 62479:2010, EN 50663:2017	CISRR24041107003	Conform
	Safety	IEC 62368-1:2018 EN IEC 62368-1:2020+A11:2020	CISSR240411018	Conform
Art. 3.1b	EMC	ETSI EN 301 489-1 V2.2.3 ETSI EN 301 489-17 V3.2.4	CISRR24041107002	Conform
Art. 3.2	Radio	ETSI EN 300 328 V2.2.2	CISRR24041107001	Conform

The test results are traceable to the international or national standards.

Laboratory: Shenzhen BangCe Testing Technology Co., Ltd.

Add.: 101, Building 10, Yun Li Intelligent Park, Guang Ming District, Shenzhen, Guangdong Province China.

Tel:86-755-2319 6848, Web: <http://www.cis-cn.net/>



Approved By: Genry Long